

EC Declaration of Conformity

Manufacturer:

DeVilbiss Healthcare LLC 100 DeVilbiss Drive Somerset, PA 15501, USA

EC Authorized Representative:

DeVilbiss Healthcare GmbH Kamenzerstraße 3, 68309 Mannheim, Germany

1. Oxygen Delivery Units, controlled (UMDNS 18-076):

Catalogue nos.:

5351

Classification (MDD Annex IX):

IIa (Rule 11)

Conformity Assessment Procedure:

MDD 93/42/EEC, Annex II excluding Section 4

2. Accessories:

Product Description (Catalogue no.):

ML6 Cylinder w/CF Regulator, Europe 5351-ML6-CF C Cylinder w/CF Regulator, Europe 5351-C-CF D Cylinder w/CF Regulator, Europe 5351-D-CF E Cylinder w/CF Regulator, Europe 535I-E-CF Cabinet Air Filter, 6 pk. 535D-605 Bag, C Cylinder EX3000D-651 Bag, D Cylinder EX3000D-652 Bag, M-6 Mini Cylinder EX3000D-653 Bag, ML6 Cylinder EX3000D-654 Cylinder Nipple Dust Cap, 5 pk. PD1000A-627 Covers, Top & Bottom replacement PD1000A-601 Gauge Kit PD1000A-620 Gauge Kit PD1000G-620 Seal, Regulator, 10 pk. PD1000G-618 PD1000A w/C Cylinder, Europe PD1000A-I-C PD1000A w/D Cylinder, Europe PD1000A-I-D PD1000A w/E Cylinder, Europe PD1000A-I-E PD1000A w/ML6 Cylinder, Europe PD1000A-I-ML6

Applied standards: All applicable harmonized standards as adopted under the EC Directive 93/42/EEC and published in the Official Journal of the European Communities. (See attached listing)

This Declaration of Conformity is based on the EC Directive 93/42/EEC, Annex II.3 and supported by the TÜV NORD CERT GmbH (0044) certificate, with reference to articles 1 and 3 of the directive 93/42/EEC.

Notified Body:

TÜV NORD CERT GmbH

Langemarckstrasse 20, 45141 Essen, Germany

Identification No.:

0044

EC Certificate No.: Start of EC Marking:

44 232 117803 2002-06-19



EC Declaration of Conformity

We, DeVilbiss Healthcare LLC., hereby declare under our sole responsibility that the above-mentioned products comply with the essential requirements and provisions of Council Directive 93/42/EEC.

Validity of this Declaration:

Somerset, PA, June 29, 2023
Place, Date

Name and Positio

Name and Position

2019-08-07 - 2024-05-26

Roberto Munoz Director, Regulatory Affairs and Audit

Applied Standards:

535 series

IEC 60601-1:2005+A1:2012 Medial electrical equipment — Part 1 General requirements for basic safety
IEC 60601-1-2:2014, Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic compatibility — Requirements and tests (FDA Recognition Number 19-8)

IEC 60601-1-6:2010 + AMD 1:2013 Ed. 3.1 Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability (associated with IEC 60601-1 Ed. 3.0) (FDA Recognition Number 5-89)

IEC 60601-1-9:2007 + A1:2013 Ed. 1.1 Medical electrical equipment - Part 1-9: General requirements for basic safety and essential performance - Collateral standard: Requirements for environmentally conscious design (associated with IEC 60601-1 Ed. 3.0)

IEC 60601-1-11:2015 (Ed 2.0), Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment (FDA Recognition Number19-14)

IEC 62366:2007 Ed. 1.0 + AMD 1:2014 — Medical devices - Application of usability engineering to medical devices (FDA Recognition number 5-87)

IEC 62304:2006 + AMD1:2015: Ed. 1.1 - Medical device software - Software life cycle processes (FDA Recognition number 13-79)

BS EN ISO 14971:2012 Medical devices - Application of risk management to medical devices (FDA Recognition number 5-40 is to ISO 14971:2007

ISTA Procedure 3A: Packaged-Products for Parcel Delivery System Shipments 70kg (150 lb) or less (standard).

ASTM G63-15 Standard Guide for Evaluating Nonmetallic Materials for Oxygen Service

ASTM G94-05 Standard Guide for Evaluating Metals for Oxygen Service

BS EN ISO 15001:2011 Anaesthetic and respiratory equipment. Compatibility with oxygen